

Detection of *mecA*- and *mecC*-Positive Methicillin-Resistant *Staphylococcus aureus* (MRSA) Isolates by the New Xpert MRSA Gen 3 PCR Assay

Karsten Becker,^a Olivier Denis,^b Sandrine Roisin,^b Alexander Mellmann,^c Evgeny A. Idelevich,^a Dennis Knaack,^a Sarah van Alen,^a André Kriegeskorte,^a Robin Köck,^a Frieder Schaumburg,^a Georg Peters,^a Britta Ballhausen^a

Institute of Medical Microbiology, University Hospital Münster, Münster, Germany^a; *Staphylococcus* National Reference Centre, Erasme Hospital, Université Libre de Bruxelles, Brussels, Belgium^b; Institute of Hygiene, University Hospital Münster, Münster, Germany^c

An advanced methicillin-resistant *Staphylococcus aureus* (MRSA) detection PCR approach targeting *SCCmec-orfX* along with *mecA* and *mecC* was evaluated for *S. aureus* and coagulase-negative staphylococci. The possession of *mecA* and/or *mecC* was correctly confirmed in all cases. All methicillin-susceptible *S. aureus* strains ($n = 98$; including staphylococcal cassette chromosome *mec* element [*SCCmec*] remnants) and 98.1% of the MRSA strains ($n = 160$, including 10 *mecC*-positive MRSA) were accurately categorized.

Rapid methicillin-resistant *Staphylococcus aureus* (MRSA) tests are based upon either the multiple-locus approach, which targets both the resistance determinant *mecA* and an *S. aureus* species-specific target, or the single-locus approach that targets the junction between the staphylococcal cassette chromosome *mec* element (*SCCmec*) and *orfX*. The high diversity of *SCCmec* and its presence in clinically relevant coagulase-negative staphylococcal (CoNS) species may lead to false-positive and/or false-negative results in both approaches, depending on the target structures (1, 2). Hence, combining the two target strategies might help overcome some of the detection and interpretation disadvantages of currently available assays (3). In addition, detection of the recently reported *mecA* homologue, designated *mecC*, should be included (4–8).

This report describes the evaluation of an advanced rapid MRSA assay that includes primers and probes for the detection of *mecA* and *mecC* along with the detection of the *SCCmec-orfX* junction. In addition to the *SCCmec* types I to IV (including subtype IVa), and V, which were already covered by the previous version, the Xpert MRSA Gen 3 assay also detects *SCCmec* types VI to XI (package insert; Xpert MRSA Gen 3, Cepheid, 2014).

Using the GeneXpert automatic system (Cepheid, Sunnyvale, CA), the Xpert MRSA Gen 3 PCR assay (Cepheid) was tested on a total of 308 isolates comprising clinical, type, and reference strains, including 17 staphylococcal species and subspecies (*S. aureus*, $n = 258$; CoNS, $n = 50$) (Table 1). At the time of the study, the Xpert MRSA Gen3 assay was designated research use only (RUO). The current regulatory status of this assay is *in vitro* diagnostic use only in the CE market. All isolates were recovered from clinical specimens during the course of several German and Belgian multicenter studies (9–16). Of these, the *mecA*-positive MRSA strains ($n = 150$) comprised the 50 most prevalent *S. aureus* protein A gene (*spa*) types found in Germany (12) (Table 1). The *mecC*-positive MRSA strains ($n = 10$) were collected in Germany and the Netherlands and exhibited six different *spa* types (6, 12, 17) (Table 1). Additionally, 98 methicillin-susceptible *S. aureus* (MSSA) isolates covering 70 *spa* types were tested, including 10 isolates known to give false-positive results in the previous version of the test (Xpert MRSA assay), and four previously deter-

mined *SCCmec* “dropout” strains, i.e., former MRSA strains that had lost major parts of the *SCCmec* element, including the *mec* genes, but still carry short remains, which might serve as a primer target in single-locus PCR approaches (9, 16) (see Table S1 in the supplemental material). Finally, 25 MR-CoNS and 25 MS-CoNS strains comprising 16 species and subspecies were tested (Table 1). Species identification, detection of *mecA* and *mecC*, and *SCCmec* typing were done as described previously (6, 18–21).

To mimic the *in vivo* situation, 1.5×10^4 bacterial cells were used from a fresh overnight culture in 100 μ l and transferred to the test cartridge. This was followed by application of the assay protocol as indicated in the Xpert MRSA Gen 3 package insert. The interpretation of the assay results and categorization as MRSA were done as recommended by the manufacturer (with “MRSA detected” meaning that both the *SCCmec-orfX* and *mecA-mecC* targets tested positive, and “MRSA not detected” meaning that one or both of the *SCCmec-orfX* and *mecA-mecC* targets tested negative). To solve discrepant results, whole-genome sequencing (WGS) was performed as recently described (22). The resulting raw reads were mapped to the *SCCmec-orfX* regions of the respective *SCCmec* type of the reference genomes (see Table S2 in the supplemental material) after quality trimming using the BWA algorithm, with default parameters, implemented in the SeqSphere⁺ software version 2.3 (Ridom, Münster, Germany).

Received 30 July 2015 Returned for modification 16 August 2015

Accepted 17 October 2015

Accepted manuscript posted online 21 October 2015

Citation Becker K, Denis O, Roisin S, Mellmann A, Idelevich EA, Knaack D, van Alen S, Kriegeskorte A, Köck R, Schaumburg F, Peters G, Ballhausen B. 2016. Detection of *mecA*- and *mecC*-positive methicillin-resistant *Staphylococcus aureus* (MRSA) isolates by the new Xpert MRSA Gen 3 PCR assay. *J Clin Microbiol* 54:180–184. doi:10.1128/JCM.02081-15.

Editor: C.-A. D. Burnham

Address correspondence to Karsten Becker, kbecker@uni-muenster.de.

Supplemental material for this article may be found at <http://dx.doi.org/10.1128/JCM.02081-15>.

Copyright © 2015, American Society for Microbiology. All Rights Reserved.

TABLE 1 Number of staphylococcal strains included and results of evaluation the Xpert MRSA Gen 3 PCR assay

Isolate (<i>n</i>) ^a	Results of Xpert MRSA Gen 3 PCR assay				Interpretation as MRSA (no. [%]) ^b
	No. (% agreement) of isolates tested				
	<i>mecA</i> and/or <i>mecC</i>		SCC <i>mec-orfX</i>		
	Positive	Negative	Positive	Negative	
MRSA					
<i>mecA</i> -positive MRSA (150) ^c	150 (100.0)	0	147 (98.0)	3 (2.0) ^d	147 (98.0)
<i>mecC</i> -positive MRSA (10) ^e	10 (100.0)	0	10 (100.0)	0	10 (100.0)
Total (160)	160 (100.0)	0	157 (98.1)	3 (1.9)	157 (98.1)
MSSA (98) ^f	0	98 (100.0)	27 (28.1)	69 (71.9)	0
MR-CoNS (25) ^g	25 (100.0)	0	0	25 (100.0)	0
MS-CoNS (25) ^g	0	25 (100.0)	0	25 (100.0)	0

^a MRSA, methicillin-resistant *S. aureus*; MSSA, methicillin-susceptible *S. aureus*; MR-CoNS, methicillin-resistant coagulase-negative staphylococci; MS-CoNS, methicillin-susceptible coagulase-negative staphylococci.

^b Interpretation of the assay results and categorization as MRSA as given by the manufacturer: MRSA detected, both SCC*mec-orfX* and *mecA* and/or *mecC* targets tested positive; MRSA not detected, one or both of the SCC*mec-orfX* and *mecA* and/or *mecC* targets tested negative.

^c Including more frequently encountered (t001, t002, t003, t004, t011, t008, t014, t020, t022, t024, t032, t034, t045, t264, t463, t1227, t2373, t4217, t4881, and t8374, each *n* = 6) and rarely occurring (t012, t015, t030, t037, t038, t041, t044, t063; t114, t127, t151, t223, t318, t379, t437, t481, t504, t535, t578, t634, t651, t785, t849, t1107, t1282, t2369, t4417, t6736, t7391, and t8380, each *n* = 1) MRSA *spa* types in Germany. SCC*mec* types I (*n* = 7), II (*n* = 57), III (*n* = 2), IV (*n* = 70) and V (*n* = 13) were detected; one isolate was nontypeable.

^d The raw reads of the three isolates (RUO83 [t004], RUO140 [t003], and RUO159 [t004]) were submitted to the European Nucleotide Archive (<http://www.ebi.ac.uk/ena/>) under the study accession no. PRJEB10686.

^e Comprising *spa* types t843 (*n* = 5), t978, t1773, t5930, t7189, and t7603.

^f Includes 4 known SCC*mec* remnant strains and 10 isolates giving false-positive results in the previous version of the Xpert MRSA Gen PCR assay.

^g Comprising 30 clinical *Staphylococcus* strains (*S. capitis* subsp. *capitis*, *n* = 1; *S. epidermidis*, *n* = 10; *S. haemolyticus*, *n* = 10; *S. hominis* subsp. *hominis*, *n* = 8; and *S. warneri*, *n* = 1) and 20 type and reference strains (*S. auricularis* DSM 20609; *S. cohnii* subsp. *cohnii* DSM 20260; *S. cohnii* subsp. *urealyticus* DSM 6718; *S. haemolyticus* DSM 20263 and DSM 20264; *S. hominis* subsp. *hominis* DSM 20320, DSM 20328, and DSM 20329; *S. hominis* subsp. *novobiosepticus* ATCC 700236; *S. hyicus* DSM 20459; *S. lugdunensis* DSM 4804, DSM 4805, and DSM 6670; *S. saprophyticus* subsp. *saprophyticus* DSM 20229 and DSM 20289; *S. schleiferi* subsp. *schleiferi* DSM 6628; *S. sciuri* subsp. *sciuri* DSM 20345; *S. simulans* DSM 20322; *S. warneri* DSM 20316; and *S. xylosus* DSM 6179).

Overall, 157/160 (98.1% positive agreement) of the MRSA strains were correctly categorized by the novel assay. While all 10 *mecC*-positive strains were detected and classified as MRSA, 3 (2.0%) of the 150 *mecA*-positive MRSA strains were falsely categorized as MSSA due to missing amplification of the SCC*mec-orfX* junction (Table 1). Another assay targeting the SCC*mec-orfX* junction (BD Max MRSA XT kit; BD Diagnostics, Quebec, Canada) likewise failed (not shown). The unequivocal possession of the *mecA* gene of these strains was reviewed by applying a further molecular assay (GenoType MRSA; Hain Lifescience, Nehren, Germany; data not shown). WGS did not show sequence variations within *orfX* compared to published sequences of respective *S. aureus* reference strains. Two of these isolates belonged to *spa* type t004 (sequence type 45 [ST45]) and carried SCC*mec* type IV. The *orfX*-SCC*mec* junction of both isolates showed a 52-kb deletion compared to *S. aureus* CA-347 beginning 326 bp downstream of *orfX*. For the third isolate (*spa* type t003, ST225, SCC*mec* type III), no considerable sequence variants were found within 3.5 kb downstream of *orfX* compared to *S. aureus* strains N315 and MW2.

All 98 MSSA strains tested showed negative results for *mecA*-*mecC* amplification (100% negative agreement) and hence were correctly categorized as MSSA (see Table S1 in the supplemental material), similar to in a recent prospective study (23). Twenty-seven MSSA strains showed positive results in SCC*mec-orfX* junction testing, including those previously characterized as SCC*mec* remnant strains, and eight of the 10 isolates tested false positive in the previous version of the Xpert MRSA PCR assay (see Table S1). The four SCC*mec* remnant strains belonged to *spa* types t011, t038, and t068 (*n* = 2). Other SCC*mec-orfX* junction-positive

MSSA strains were characterized by *spa* types t002, t008, t216, t364, t369, t5160, and t6752 or belonged to the *spa* clonal complex 127 (*spa*-CC127) (t127, t177, and t948) (see Table S1). The assay detected the *mecA* and *mecC* genes correctly in all CoNS.

The detection of *mecC*-positive MRSA is a major advantage of the Xpert MRSA Gen 3 PCR assay. Diagnostics were challenged by the recent discovery of the *mecA*_{LGA251} (*mecC*) gene as part of a novel SCC*mec* XI element in *S. aureus* (4, 5) and CoNS (24–26). With the spread of *mecC*-harboring MRSA (4–7, 12, 27–32), the absence of the *mecA* gene alone can no longer be considered a reliable genetic marker to exclude MRSA. The failure in conventional *mecA* detection assays to detect *mecC* results in inconsistent results in comparison to those with phenotypic susceptibility tests (4–6). Besides various in-house PCR procedures (4–6, 27, 33), another commercially available multiplex PCR, based upon a multiple-locus detection strategy, was recently shown to be able to detect *mecC*-positive MRSA (10). Moreover, both the genetic diversity of the strain background and the occurrence of *mecC*-harboring staphylococci in livestock, wildlife, and environmental sources are worrisome (7, 34–39).

The combined detection of *mecA* and *mecC* with the SCC*mec-orfX* junction represents a second major advantage. It overcomes the problem due to *mecA*- and *mecC*-negative remnants of the SCC*mec* element, which may cause false-positive results (16, 40–44). Outbreaks with SCC*mec* remnant MSSA isolates may result in medical and economic burden due to unjustified MRSA precaution measures (45). The inclusion of the *mecA* and *mecC* genes as targets overcomes this source of misinterpretation. All four SCC*mec* remnant strains included were categorized as MSSA.

In the case of the cooccurrence of an MR-CoNS and an

SCC*mec* remnant MSSA in clinical specimens, false-positive results may still arise. Here, the inclusion of another *S. aureus*-specific target gene sequence might clarify this problem (46, 47). The detection of *mecA*-*mecC* amplification along with a negative result for the SCC*mec*-*orfX* junction will be categorized as MRSA not detected, according to the manufacturer's instructions. In this case, the presence of MR-CoNS could be assumed. However, in rare cases, this diagnostic pattern might also indicate a false-negative result if unknown or uncovered nucleic acid variations in the *orfX* region-neighboring part of the SCC*mec* elements hampered the correct identification (48–52). Here, three *mecA*-positive strains were not detected. Two of these strains harbored a deletion close downstream of *orfX* that might explain the failure by a possible loss of the respective primer-binding site; the reason for misidentification of the other strain, determined by another junction-targeting PCR approach, remains unknown. Those strains could remain undetected for a long time, thus necessitating constant monitoring of the local MRSA epidemiology (52–54).

In conclusion, the inclusion of *mecA* and *mecC* as targets closed a gap in the molecular detection of MRSA and minimized the risk of false-positive interpretation as MRSA due to SCC*mec* remnant isolates. The evaluated MRSA assay challenged by a large collection of German and Belgium clonal MRSA lineages was able to detect the *mecA* and *mecC* genes, respectively, of all strains included and correctly categorized the vast majority of MRSA and all non-MRSA strains.

Nucleotide sequence accession number. The raw reads of the three isolates (RUO83 [t004], RUO140 [t003], and RUO159 [t004]) were submitted to the European Nucleotide Archive (<http://www.ebi.ac.uk/ena/>) under study accession no. PRJEB10686.

ACKNOWLEDGMENTS

We thank M. Schulte and A. Hassing for excellent technical assistance. This work was supported in part by a grant from BMBF (MedVet-Staph) to K.B., R.K., and A.M. (grant 01KI1301A).

K.B. has received lecture fees from Cepheid.

FUNDING INFORMATION

Bundesministerium für Bildung und Forschung (Federal Ministry of Education and Research) provided funding to Karsten Becker, Alexander Mellmann, and Robin Köck under grant number 01KI1301A.

The Xpert MRSA Gen 3 compounds were provided free of charge by Cepheid.

REFERENCES

- Becker K, Heilmann C, Peters G. 2014. Coagulase-negative staphylococci. *Clin Microbiol Rev* 27:870–926. <http://dx.doi.org/10.1128/CMR.00109-13>.
- Shore AC, Coleman DC. 2013. Staphylococcal cassette chromosome *mec*: recent advances and new insights. *Int J Med Microbiol* 303:350–359. <http://dx.doi.org/10.1016/j.ijmm.2013.02.002>.
- Nijhuis RH, van Maarseveen NM, van Hannen EJ, van Zwet AA, Mascini EM. 2014. A rapid and high-throughput screening approach for methicillin-resistant *Staphylococcus aureus* based on the combination of two different real-time PCR assays. *J Clin Microbiol* 52:2861–2867. <http://dx.doi.org/10.1128/JCM.00808-14>.
- García-Álvarez M, Holden MT, Lindsay H, Webb CR, Brown DF, Curran MD, Walpole E, Brooks K, Pickard DJ, Teale C, Parkhill J, Bentley SD, Edwards GF, Girvan EK, Kearns AM, Pichon B, Hill RL, Larsen AR, Skov RL, Peacock SJ, Maskell DJ, Holmes MA. 2011. Methicillin-resistant *Staphylococcus aureus* with a novel *mecA* homologue in human and bovine populations in the UK and Denmark: a descriptive study. *Lancet Infect Dis* 11:595–603. [http://dx.doi.org/10.1016/S1473-3099\(11\)70126-8](http://dx.doi.org/10.1016/S1473-3099(11)70126-8).
- Shore AC, Deasy EC, Slickers P, Brennan G, O'Connell B, Monecke S, Ehrlich R, Coleman DC. 2011. Detection of staphylococcal cassette chromosome *mec* type XI encoding highly divergent *mecA*, *mecI*, *mecR1*, *blaZ* and *ccr* genes in human clinical clonal complex 130 methicillin-resistant *Staphylococcus aureus*. *Antimicrob Agents Chemother* 55:3765–3773. <http://dx.doi.org/10.1128/AAC.00187-11>.
- Kriegeskorte A, Ballhausen B, Idelevich EA, Köck R, Friedrich AW, Karch H, Peters G, Becker K. 2012. Human MRSA isolates with novel genetic homolog, Germany. *Emerg Infect Dis* 18:1016–1018. <http://dx.doi.org/10.3201/eid1806.110910>.
- Becker K, Ballhausen B, Köck R, Kriegeskorte A. 2014. Methicillin resistance in *Staphylococcus* isolates: the “*mec* alphabet” with specific consideration of *mecC*, a *mec* homolog associated with zoonotic *S. aureus* lineages. *Int J Med Microbiol* 304:794–804. <http://dx.doi.org/10.1016/j.ijmm.2014.06.007>.
- Ballhausen B, Kriegeskorte A, Schleimer N, Peters G, Becker K. 2014. The *mecA* homolog *mecC* confers resistance against β -lactams in *Staphylococcus aureus* irrespective of the genetic strain background. *Antimicrob Agents Chemother* 58:3791–3798. <http://dx.doi.org/10.1128/AAC.02731-13>.
- Roisin S, Laurent C, Nonhoff C, Deplano A, Hallin M, Byl B, Struelens MJ, Denis O. 2012. Positive predictive value of the Xpert MRSA assay diagnostic for universal patient screening at hospital admission: influence of the local ecology. *Eur J Clin Microbiol Infect Dis* 31:873–880. <http://dx.doi.org/10.1007/s10096-011-1387-7>.
- Becker K, Larsen AR, Skov RL, Paterson GK, Holmes MA, Sabat AJ, Friedrich AW, Köck R, Peters G, Kriegeskorte A. 2013. Evaluation of a modular multiplex-PCR methicillin-resistant *Staphylococcus aureus* (MRSA) detection assay adapted for *mecC* detection. *J Clin Microbiol* 51:1917–1919. <http://dx.doi.org/10.1128/JCM.00075-13>.
- von Eiff C, Reinert RR, Kresken M, Brauers J, Hafner D, Peters G. 2000. Nationwide German multicenter study on prevalence of antibiotic resistance in staphylococcal bloodstream isolates and comparative *in vitro* activities of quinupristin-dalfopristin. *J Clin Microbiol* 38:2819–2823.
- Schaumburg F, Köck R, Mellmann A, Richter L, Hasenberg F, Kriegeskorte A, Friedrich AW, Gatermann S, Peters G, von Eiff C, Becker K, Study Group. 2012. Population dynamics among methicillin resistant *Staphylococcus aureus* in Germany during a 6-year period. *J Clin Microbiol* 50:3186–3192. <http://dx.doi.org/10.1128/JCM.01174-12>.
- Becker K, Haverkämper G, von Eiff C, Roth R, Peters G. 2001. Survey of staphylococcal enterotoxin genes, exfoliative toxin genes, and toxic shock syndrome toxin 1 gene in non-*Staphylococcus aureus* species. *Eur J Clin Microbiol Infect Dis* 20:407–409. <http://dx.doi.org/10.1007/PL00011281>.
- Köck R, Werner P, Friedrich AW, Fegeler C, Becker K, Prevalence of Multiresistant Microorganisms (PMM) Study Group. 2012. Characteristics of *Staphylococcus aureus* nasal carriage, resistance patterns and genetic lineages in healthy German adults, abstr C2-1382. 52nd Intersci Conf Antimicrob Agents Chemother (ICAAC), 9 to 12 September 2012, San Francisco, CA.
- Vandendriessche S, Hallin M, Cattri B, Jans B, Deplano A, Nonhoff C, Roisin S, De Mendonça R, Struelens MJ, Denis O. 2012. Previous healthcare exposure is the main antecedent for methicillin-resistant *Staphylococcus aureus* carriage on hospital admission in Belgium. *Eur J Clin Microbiol Infect Dis* 31:2283–2292. <http://dx.doi.org/10.1007/s10096-012-1567-0>.
- Vandendriessche S, Vanderhaeghen W, Larsen J, de Mendonça R, Hallin M, Butaye P, Hermans K, Haesebrouck F, Denis O. 2014. High genetic diversity of methicillin-susceptible *Staphylococcus aureus* (MSSA) from humans and animals on livestock farms and presence of SCC*mec* remnant DNA in MSSA CC398. *J Antimicrob Chemother* 69:355–362. <http://dx.doi.org/10.1093/jac/dkt366>.
- Sabat AJ, Koksals M, Akkerboom V, Monecke S, Kriegeskorte A, Hendrix R, Ehrlich R, Köck R, Becker K, Friedrich AW. 2012. Detection of new methicillin-resistant *Staphylococcus aureus* strains that carry a novel genetic homologue and important virulence determinants. *J Clin Microbiol* 50:3374–3377. <http://dx.doi.org/10.1128/JCM.01121-12>.
- Mellmann A, Becker K, von Eiff C, Keckevoet U, Schumann P, Harmsen D. 2006. Sequencing and staphylococci identification. *Emerg Infect Dis* 12:333–336. <http://dx.doi.org/10.3201/eid1202.050962>.
- Becker K, Harmsen D, Mellmann A, Meier C, Schumann P, Peters G, von Eiff C. 2004. Development and evaluation of a quality-controlled ribosomal sequence database for 16S ribosomal DNA-based identification

- of *Staphylococcus* species. *J Clin Microbiol* 42:4988–4995. <http://dx.doi.org/10.1128/JCM.42.11.4988-4995.2004>.
20. Becker K, Pagnier I, Schuhen B, Wenzelburger F, Friedrich AW, Kipp F, Peters G, von Eiff C. 2006. Does nasal cocolonization by methicillin-resistant coagulase-negative staphylococci and methicillin-susceptible *Staphylococcus aureus* strains occur frequently enough to represent a risk of false-positive methicillin-resistant *S. aureus* determinations by molecular methods? *J Clin Microbiol* 44:229–231. <http://dx.doi.org/10.1128/JCM.44.1.229-231.2006>.
 21. Okuma K, Iwakawa K, Turnidge JD, Grubb WB, Bell JM, O'Brien FG, Coombs GW, Pearman JW, Tenover FC, Kapi M, Tiensasitorn C, Ito T, Hiramatsu K. 2002. Dissemination of new methicillin-resistant *Staphylococcus aureus* clones in the community. *J Clin Microbiol* 40:4289–4294. <http://dx.doi.org/10.1128/JCM.40.11.4289-4294.2002>.
 22. Ruppitsch W, Pietzka A, Prior K, Bletz S, Fernandez HL, Allerberger F, Harmsen D, Mellmann A. 2015. Defining and evaluating a core genome MLST scheme for whole genome sequence-based typing of *Listeria monocytogenes*. *J Clin Microbiol*, in press. <http://dx.doi.org/10.1128/JCM.01193-15>.
 23. Lepointeur M, Delattre S, Cozza S, Lawrence C, Roux AL, Rottman M. 2015. Comparative evaluation of two PCR-based methods for detection of methicillin-resistant *Staphylococcus aureus* (MRSA): Xpert MRSA Gen 3 and BD-Max MRSA XT. *J Clin Microbiol* 53:1955–1958. <http://dx.doi.org/10.1128/JCM.03679-14>.
 24. Loncaric I, Küber-Heiss A, Posautz A, Stalder GL, Hoffmann D, Rosengarten R, Walzer C. 2013. Characterization of methicillin-resistant *Staphylococcus* spp. carrying the *mecC* gene, isolated from wildlife. *J Antimicrob Chemother* 68:2222–2225. <http://dx.doi.org/10.1093/jac/dkt186>.
 25. Harrison EM, Paterson GK, Holden MT, Morgan FJ, Larsen AR, Petersen A, Leroy S, De Vliegher S, Perreten V, Fox LK, Lam TJ, Sampimon OC, Zadoks RN, Peacock SJ, Parkhill J, Holmes MA. 2013. A *Staphylococcus xylosum* isolate with a new *mecC* allotype. *Antimicrob Agents Chemother* 57:1524–1528. <http://dx.doi.org/10.1128/AAC.01882-12>.
 26. Harrison EM, Paterson GK, Holden MT, Ba X, Rolo J, Morgan FJ, Pichon B, Kearns A, Zadoks RN, Peacock SJ, Parkhill J, Holmes MA. 2014. A novel hybrid SCCmec-mecC region in *Staphylococcus sciuri*. *J Antimicrob Chemother* 69:911–918. <http://dx.doi.org/10.1093/jac/dkt452>.
 27. Cuny C, Leyer F, Strommenger B, Witte W. 2011. Rare occurrence of methicillin-resistant *Staphylococcus aureus* CC130 with a novel *mecA* homologue in humans in Germany. *PLoS One* 6:e24360. <http://dx.doi.org/10.1371/journal.pone.0024360>.
 28. Laurent F, Chardon H, Haenni M, Bes M, Reverdy ME, Madec JY, Lagier E, Vandenesch F, Tristan A. 2012. MRSA harboring *mecA* variant gene *mecC*, France. *Emerg Infect Dis* 18:1465–1467. <http://dx.doi.org/10.3201/eid1809.111920>.
 29. Medhus A, Slettemeås JS, Marstein L, Larssen KW, Sunde M. 2012. Methicillin-resistant *Staphylococcus aureus* with the novel *mecC* gene variant isolated from a cat suffering from chronic conjunctivitis. *J Antimicrob Chemother* 68:968–969. <http://dx.doi.org/10.1093/jac/dks487>.
 30. Paterson GK, Harrison EM, Holmes MA. 2014. The emergence of *mecC* methicillin-resistant *Staphylococcus aureus*. *Trends Microbiol* 22:42–47. <http://dx.doi.org/10.1016/j.tim.2013.11.003>.
 31. Petersen A, Stegger M, Helberg O, Christensen J, Zeuthen A, Knudsen LK, Urth T, Sorum M, Schouls L, Larsen J, Skov R, Larsen AR. 2012. Epidemiology of methicillin-resistant *Staphylococcus aureus* carrying the novel *mecC* gene in Denmark corroborates a zoonotic reservoir with transmission to humans. *Clin Microbiol Infect* 19:E16–E22. <http://dx.doi.org/10.1111/1469-0691.12036>.
 32. Deplano A, Vandendriessche S, Nonhoff C, Denis O. 2014. Genetic diversity among methicillin-resistant *Staphylococcus aureus* isolates carrying the *mecC* gene in Belgium. *J Antimicrob Chemother* 69:1457–1460. <http://dx.doi.org/10.1093/jac/dku020>.
 33. Stegger M, Andersen PS, Kearns A, Pichon B, Holmes MA, Edwards G, Laurent F, Teale C, Skov R, Larsen AR. 2012. Rapid detection, differentiation and typing of methicillin-resistant *Staphylococcus aureus* harbouring either *mecA* or the new *mecA* homologue *mecA_{LGA251}*. *Clin Microbiol Infect* 18:395–400. <http://dx.doi.org/10.1111/j.1469-0691.2011.03715.x>.
 34. Monecke S, Gavier-Widen D, Mattsson R, Rangstrup-Christensen L, Lazaris A, Coleman DC, Shore AC, Ehrlich R. 2013. Detection of *mecC*-positive *Staphylococcus aureus* (CC130-MRSA-XI) in diseased European hedgehogs (*Erinaceus europaeus*) in Sweden. *PLoS One* 8:e66166. <http://dx.doi.org/10.1371/journal.pone.0066166>.
 35. Loncaric I, Küber-Heiss A, Posautz A, Stalder GL, Hoffmann D, Rosengarten R, Walzer C. 2014. *mecC*- and *mecA*-positive methicillin-resistant *Staphylococcus aureus* (MRSA) isolated from livestock sharing habitat with wildlife previously tested positive for *mecC*-positive MRSA. *Vet Dermatol* 25:147–148. <http://dx.doi.org/10.1111/vde.12116>.
 36. Porrero MC, Valverde A, Fernández-Llario P, Díez-Guerrier A, Mateos A, Lavín S, Cantón R, Fernández-Garayzabal JF, Domínguez L. 2014. *Staphylococcus aureus* carrying *mecC* gene in animals and urban wastewater, Spain. *Emerg Infect Dis* 20:899–901. <http://dx.doi.org/10.3201/eid2005.130426>.
 37. Schlotter K, Huber-Schlenstedt R, Gangl A, Hotzel H, Monecke S, Müller E, Reißig A, Proft S, Ehrlich R. 2014. Multiple cases of methicillin-resistant CC130 *Staphylococcus aureus* harboring *mecC* in milk and swab samples from a Bavarian dairy herd. *J Dairy Sci* 97:2782–2788. <http://dx.doi.org/10.3168/jds.2013-7378>.
 38. Paterson GK, Morgan FJ, Harrison EM, Peacock SJ, Parkhill J, Zadoks RN, Holmes MA. 2014. Prevalence and properties of *mecC* methicillin-resistant *Staphylococcus aureus* (MRSA) in bovine bulk tank milk in Great Britain. *J Antimicrob Chemother* 69:598–602. <http://dx.doi.org/10.1093/jac/dkt417>.
 39. Eriksson J, Espinosa-Gongora C, Stamphøj I, Larsen AR, Guardabassi L. 2013. Carriage frequency, diversity and methicillin resistance of *Staphylococcus aureus* in Danish small ruminants. *Vet Microbiol* 163:110–115. <http://dx.doi.org/10.1016/j.vetmic.2012.12.006>.
 40. Shore AC, Rossney AS, O'Connell B, Herra CM, Sullivan DJ, Humphreys H, Coleman DC. 2008. Detection of staphylococcal cassette chromosome *mec*-associated DNA segments in multiresistant methicillin-susceptible *Staphylococcus aureus* (MSSA) and identification of *Staphylococcus epidermidis ccrAB4* in both methicillin-resistant *S. aureus* and MSSA. *Antimicrob Agents Chemother* 52:4407–4419. <http://dx.doi.org/10.1128/AAC.00447-08>.
 41. Stojanov M, Blanc DS. 2014. Characterization of the staphylococcal cassette chromosome *mec* insertion site in 108 isolates lacking the *mecA* gene and identified as methicillin-resistant *Staphylococcus aureus* by the Xpert MRSA assay. *Eur J Clin Microbiol Infect Dis* 33:1967–1971. <http://dx.doi.org/10.1007/s10096-014-2169-9>.
 42. Arbefeville SS, Zhang K, Kroeger JS, Howard WJ, Diekema DJ, Richter SS. 2011. Prevalence and genetic relatedness of methicillin-susceptible *Staphylococcus aureus* isolates detected by the Xpert MRSA nasal assay. *J Clin Microbiol* 49:2996–2999. <http://dx.doi.org/10.1128/JCM.00046-11>.
 43. Blanc DS, Bassat P, Nahimana-Tessemo I, Jaton K, Greub G, Zanetti G. 2011. High proportion of wrongly identified methicillin-resistant *Staphylococcus aureus* carriers by use of a rapid commercial PCR assay due to presence of staphylococcal cassette chromosome element lacking the *mecA* gene. *J Clin Microbiol* 49:722–724. <http://dx.doi.org/10.1128/JCM.01988-10>.
 44. Edwards AM, Massey RC, Clarke SR. 2012. Molecular mechanisms of *Staphylococcus aureus* nasopharyngeal colonization. *Mol Oral Microbiol* 27:1–10. <http://dx.doi.org/10.1111/j.2041-1014.2011.00628.x>.
 45. Lindqvist M, Isaksson B, Grub C, Jonassen TØ, Hällgren A. 2012. Detection and characterisation of SCCmec remnants in multiresistant methicillin-susceptible *Staphylococcus aureus* causing a clonal outbreak in a Swedish county. *Eur J Clin Microbiol Infect Dis* 31:141–147. <http://dx.doi.org/10.1007/s10096-011-1286-y>.
 46. Hussain M, von Eiff C, Sinha B, Joost I, Herrmann M, Peters G, Becker K. 2008. *eap* gene as novel target for specific identification of *Staphylococcus aureus*. *J Clin Microbiol* 46:470–476. <http://dx.doi.org/10.1128/JCM.01425-07>.
 47. Brakstad OG, Aasbakk K, Maeland JA. 1992. Detection of *Staphylococcus aureus* by polymerase chain reaction amplification of the *nuc* gene. *J Clin Microbiol* 30:1654–1660.
 48. Ito T, Ma XX, Takeuchi F, Okuma K, Yuzawa H, Hiramatsu K. 2004. Novel type V staphylococcal cassette chromosome *mec* driven by a novel cassette chromosome recombinase, *ccrC*. *Antimicrob Agents Chemother* 48:2637–2651. <http://dx.doi.org/10.1128/AAC.48.7.2637-2651.2004>.
 49. Zhang K, McClure JA, Elsayed S, Conly JM. 2009. Novel staphylococcal cassette chromosome *mec* type, tentatively designated type VIII, harboring class A *mec* and type 4 *ccr* gene complexes in a Canadian epidemic strain of methicillin-resistant *Staphylococcus aureus*. *Antimicrob Agents Chemother* 53:531–540. <http://dx.doi.org/10.1128/AAC.01118-08>.
 50. Oliveira DC, Milheiro C, de Lencastre H. 2006. Redefining a structural variant of staphylococcal cassette chromosome *mec*, SCCmec type VI. An-

- timicrob Agents Chemother 50:3457–3459. <http://dx.doi.org/10.1128/AAC.00629-06>.
51. Huletsky A, Giroux R, Rossbach V, Gagnon M, Vaillancourt M, Bernier M, Gagnon F, Truchon K, Bastien M, Picard FJ, van Belkum A, Ouellette M, Roy PH, Bergeron MG. 2004. New real-time PCR assay for rapid detection of methicillin-resistant *Staphylococcus aureus* directly from specimens containing a mixture of staphylococci. *J Clin Microbiol* 42:1875–1884. <http://dx.doi.org/10.1128/JCM.42.5.1875-1884.2004>.
52. Bartels MD, Boye K, Rohde SM, Larsen AR, Torfs H, Bouchy P, Skov R, Westh H. 2009. A common variant of staphylococcal cassette chromosome *mec* type IVa in isolates from Copenhagen, Denmark, is not detected by the BD GeneOhm methicillin-resistant *Staphylococcus aureus* assay. *J Clin Microbiol* 47:1524–1527. <http://dx.doi.org/10.1128/JCM.02153-08>.
53. Hill-Cawthorne GA, Hudson LO, El Ghany MF, Piepenburg O, Nair M, Dodgson A, Forrest MS, Clark TG, Pain A. 2014. Recombinations in staphylococcal cassette chromosome *mec* elements compromise the molecular detection of methicillin resistance in *Staphylococcus aureus*. *PLoS One* 9:e101419. <http://dx.doi.org/10.1371/journal.pone.0101419>.
54. Francois P, Bento M, Renzi G, Harbarth S, Pittet D, Schrenzel J. 2007. Evaluation of three molecular assays for rapid identification of methicillin-resistant *Staphylococcus aureus*. *J Clin Microbiol* 45:2011–2013. <http://dx.doi.org/10.1128/JCM.00232-07>.